

Product:	Neolution®Air	MADE IN GERMANY
REF:	256	
Manufacturer:	DACH Schutzbekleidung GmbH & Co. KG	

<ul> <li>CAT III Personal Protective Equipment accordingly Regulation (EU) 2016/425</li> <li>Class I Medical Device according to Regulation (EU) 2017/745.</li> <li>Helmke Drum Test accordingly IEST-PR-CC003.4: Category I         We recommend the use in clean rooms accordingly EN ISO 14644-1:         ISO class 3 to ISO class 9.</li> </ul>
Particle filtering half mask FFP3 NR D accordingly EN 149:2001+A1:2009.
Additionally complies with the requirements for Type IIR medical face masks according to EN 14683:2019+AC:2019 in terms of bacterial filtration efficiency, splash resistance pressure and microbial cleanliness.
Very good skin compatibily scientifically proven by testing accordingly EN ISO 10993-5 (cytotoxicity), EN ISO 10993-10 (sensitization) and EN ISO 10993-23 (irritation): The materials used are not cytotoxic, not sensitizing and not irritating.
<ul> <li>Mask body: Polypropylene nonwoven</li> <li>Head straps: textile elastic made of polyester and elastane, free from natural latex and PVC</li> <li>Nose clip: metal wire covered with plastic</li> </ul>
The three-part respirator Neolution®Air impresses with high-quality materials that guarantee the wearer maximum comfort. High-tech filter material generates a high filtering performance against solid and liquid aerosols with extremely low breathing resistance.
Due to the three movable parts, the mask follows the jaw movements during speech. This ensures a constant tight fit.
Textile head straps ensure a secure fit of the mask to the face. The straps fit comfortably on the head without pinching or pulling hair.
Maximum protection even under extreme conditions: for disposable masks voluntary dolomite dust loading test passed!

WE KEEP YOU **SAFE**°

V20250909

Rotackerstraße 21 | 76437 Rastatt | Germany Tel.+49 7222 15989-0

-1|4-

info@dach-germany.de



Product performance:		Product performance according to EN 149:2001+A1:2009  Filter efficiency						
	Class							
	Class		Requirement					
		NaCl 95 I/	Maximum penetration		Actual penetration  NaCl 95 I/min Paraffin oil 95 I/min			
		min			Naci 95 I/IIIII	Paraffin oil 95 I/min		
	FFP3 NR D	1% 1%			0,07 %	0,1 %		
	Breathi	Breathing resistance						
	Class	Requirement  Maximum breathing resistance (mbar)			Test result			
					Actual breathing resistance (mbar)			
		Inhalation	Exhalation		Inhalation		Exhalation	
		30 l/min	95 I/min	160 l/min	30 l/min	95 l/min	160 l/min	
	FFP3 NR D	1,0	3,0	3,0	0,3	1,0	1,6	
	Leakage test							
	Class	Requirement			Test result			
		Maximum total inward leakage (%)		Minimum Protection factor (Fit Factor)	Total inward leakage (%) <sup>1</sup>		Protection factor (Fit Factor) <sup>1,2</sup>	
	FFP3 NR D	2		30	0,91		109,9	
	under la leakage (protec	<sup>1</sup> The leakage and protection factor values were determined as mean values with 10 test persons under laboratory conditions and are for reference only. The data does not represent the actual leakage of a person at the workplace and does not replace an individual fit test. <sup>2</sup> The fit factor (protection factor) indicates the ratio between the particle concentration outside the mask and the particle concentration inside the mask.						
	Carbon	Carbon dioxide content of the inhaled air						
	Class	Requirement (max.)		Test result				

WE KEEP YOU **SAFE**°

dach-germany.de

V20250909

-2|4-



Colour: Packaging and variants:	White Packaging Quantity						
	≥ 1,0 µr ≥ 5,0 µ	m - < 5,0 μm  m	0,00058				
		m - <1,0 μm	0,00066			18 C + /- 2 C, 50% - 10% RH	
	≥ 0,3 µ	m - < 0,5 μm				IEST-PR-CC003.4 Climatic conditions: 18°C +/- 2°C,	
	Particle size Particle count (per cm²) Helmke-Drum Test						
	Test results Helmke-Drum Test IEST-PR-CC003.4 (Average)						
	*CFU = 0	*CFU = colony forming units.		(30)			
	Type IIR		Maximum CFU/g value		Actual CFU/g value		
			Requirement		Test result		
	Microbial cleanliness						
			≥16,0 kPA		16,0 kPA		
			Requireme	nt	Test result		
	Splash resistance pressure						
	Type IIR		≥98 %		>99,9 %		
	Class Requireme		Requireme	nt	Test result		
	Bacterial filtration efficiency (BFE)						
	Product	oduct performance according to EN 14683:2019 +AC:2019			 NC:2019		
	FFP3 NR D	1%		0,58 %	0,58 %		

WE KEEP YOU **SAFE**°

-3 | 4 - V20250909



	Packaging	Quantity	
	Bag	3 pcs.	
	Shipping carton	270 pcs.	
GTIN:	256	4049825006589	
Use:	For single use. Read the instructions for use before use.		
Storage:	Store in a dry place without direct sunlight in the original packaging (see packaging).  If the storage conditions are observed, the product has a storage life of 5 years (see labeling on the product).		
Environmental sustainability and disposal:	In the case of contaminated products, the type and extent of contamination determines the disposal, and the applicable laws and regulations of the relevant country must be followed.  A non-contaminated product can be thermally recycled or disposed of in landfills without releasing toxic substances.		

-4|4-